

Myolift Conductivity Meter Does Not Register all bars

The conductivity meter of the myolift devices either shows a solid bar or it flashes. The machine is calibrated for their accessories and may or may not display correctly using AXION accessories.

However, our testing, and I did a video on this, shows that current is being delivered even if the bars are flashing. So, if your mask is soaked in a saline type solution and used under our mask, or if you are using our probes, current is still being delivered.

For optimum results, use our AXION Wave 0.3Hz, True Microcurrent.

Online URL: <https://posturepro.phpkb.cloud/article.php?id=169>