

Can I use Microcurrent with Dental Implants

True microcurrent, in the therapeutic range of under 1000 microamps, does not generate heat and is therefore safe to use around metal implants, including dental implants.

Online URL: <https://posturepro.phpkb.cloud/article.php?id=170>